



Potency Results

Sample Name: *Chill Buda THCA Preroll*

Client: Hemp Nova / Edge Off Wholesale

Client Batch ID:

Pinnacle-Analytics.com

3549 Lear Way, Suite 101

Medford OR 97504

P:(541)300-8217

Sample ID: rC-H-276-D981

Matrix: Flower

Prep Analyst: Megan E.

Analysis Method: 0630322+1 H4 4-21-2022 #1.lcm

Sampling Method: N/A

Reference Method: JCB 2009: HPLC/DAD

Analysis Batch: 8-3-2023 H4 109, 205, 247, 276, 305, 372 Flower


Date Sampled: 8/3/2023

Date Reported: 8/4/2023

Client License: N/A

For R&D Purposes Only

Total THC (THCA*0.877+d9-THC) **23.3%**
Total CBD (CBDA*0.877+CBD) **0.47%**
Moisture Content **9.4%**



Cannabinoid	% Weight	mg/g
CBDVA	<LOQ	<LOQ
CBDV	<LOQ	<LOQ
CBDA*	0.415	4.15
CBGA	0.636	6.36
CBG	<LOQ	<LOQ
CBD*	0.105	1.05
THCV	<LOQ	<LOQ
CBN	<LOQ	<LOQ
d9-THC*	0.281	2.81
d8-THC*	0.153	1.53
CBC	<LOQ	<LOQ
THCA*	26.3	263.0
Total Cannabinoids	27.89	278.9

*ORELAP Accredited Analyte

Limit Of Quantitation: 0.1%, analyte not measured

CBDA*	d9-THC*
CBGA	d8-THC*
CBD*	THCA*



These test results may not be altered or reproduced except in full without the permission of Pinnacle Analytics. These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152
 Report generated by Routine_Potency_Rev13_8-1-2023

Kris Ford, PhD
Lab Director



Quality Control Results

Analyst: Megan E.

Pinnacle-Analytics.com
3549 Lear Way, Suite 101
Medford OR 97504

Analysis Batch: 8-3-2023 H4 109, 205, 247, 276, 305, 372 Flower P:(541)300-8217

	Duplicate RPD		LCS % Recovery		Method Blank	
	H-0-D974-b	Limit	C-FL-080323	Limits	C-FB-080323	Limit
CBDA	2.97%	30%	97.9%	90-110%	<LOQ/2	LOQ/2
CBD	1.24%	30%	102.0%	90-110%	<LOQ/2	LOQ/2
d9-THC	2.04%	30%	102.0%	90-110%	<LOQ/2	LOQ/2
d8-THC	<LOQ%	30%	103.0%	90-110%	<LOQ/2	LOQ/2
THCA	1.48%	30%	92.9%	90-110%	<LOQ/2	LOQ/2


RPD: Relative Percent Difference between unknown sample and its duplicate

LCS: Laboratory Control Sample with known concentration

Case Comments: There were no divergences from ordinary Quality Control procedures or SOPs.



These test results may not be altered or reproduced except in full without the permission of Pinnacle Analytics. These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152
Report generated by Routine_Potency_Rev13_8-1-2023


Kris Ford, PhD
Lab Director